

Detection of Sleep Apnea and Snoring From ECG Signals By Tunable Q-Factor Wavelet Transformation

Dr. Vijay Kumar Garg
Associate Professor, CSE
Lovely Professional University
Jalandhar, Punjab
vijay.garg@lpu.co.in

ABSTRACT

Sleep apnea and snoring are two sleep disorders that are likely to be genuine. Sleep apnea is defined as pauses during breathing or rare breathing during sleep and snoring is a sound delivered during sleep mode due to blocked development of air in the midst of breathing. The new standard diagnostic technique for these sleep disorders is costly polysomnography that requires human experts and is done in unique laboratories. Subsequently, to detect such types of disorders, a more comfortable and less expensive technique is needed. To expose them, late researchers concentrated on signal processing and pattern recognition as alternate modes. Through applying Tunable Q-Factor wavelet transform (TQWT), the flowing research focuses on the identification of sleep apnea and snoring using ECG signals. The results obtained showed a high degree of accuracy, about 85%.

Keywords- ECG Signals, Sleep Apnea, Snoring, TQWT

1. INTRODUCTION

Sleep is a naturally occurring state where a man has lost its aggregate thought against the external condition [1] and all muscles are in quiet state alongside other erotic developments. A person travels through many stages during its sleeping hours, for example, Wake, non-rapid eye movement (NREM) and rapid eye movement (REM). Then again, sooner or later this sleep is prevented on account of some embarrassing practices, mental conditions related with it, generally called sleep disorders that further hack down the idea of rest including napping hours. A various intelligent techniques, methods and frameworks have been adopted by analysts to detect sleep and sleeping patterns along with linked sleep disorders which creates some sort of unsettling influences in sleep [4-7].

In this paper, two sleep disorders are inspected like sleep apnea and snoring, in which sleep apnea [2] has an imperative perspective for over the best daytime drowsiness. It is extremely natural, in which a person experiences disturbed breathing while sleeping [3]. This disorder builds the dangers of Type-2 diabetes, cardiovascular infections and additionally death rate from road accidents and so on. A large part of the cases of sleep apnea go undetected due to the reason for its working cost, bother and inaccessibility of analysis and testing machines. Along these lines, there are numerous strategies created which utilizes the features of ECG

signals as if it is one of the efficient and easier technology used for recognition and detection of sleep apnea. Song et al [13] used temporary ECG signals dependence and a discriminative hidden markov model for PC helped findings from OSA. Chen et al [12] set up an OSA seriousness record and used SVM for PC to help distinguishing evidence from sleep apnea. Single lead ECG signals had been used as a piece of this work. Azarbarzin [8] computed zero-intersection rate and pinnacle recurrence from snoring sound signals and performed arrangement utilizing LDA. Schlotthauer et al [9] performed observational mode disintegration of heartbeat oximetry signals for computerized OSAS distinguishing proof. Hassan et al [11] used spectral features in the wavelet transform domain and used bootstrap collection for automated sleep detection in the signal domain [10] and in the double tree complex. In snoring, a sound is passed on because of unsoiled air change amidst breathing during sleeping time. Smolen et al [14] proposed the estimation procedure and sleep scoring gadget model of a home care in which nature of sleep was evaluated from video-recording subjects movement. Sound registered acoustic impacts and the main electrical flag recorded from the body surface was the signal-led ECG. In [14], it was discovered that through all phases of sleep, light snorers wheezed regularly. The most interesting fact was that in the rapid eye development sleep phase major snorers appear to wheeze more with the greatest wheezing energy than in some other sleeping stages.

All, the above said sleep disorders are sufficient fit to obstruct the physical, psychological, cognitive and motor functionality of human body. With the movement of time, these sleep disorders end up being particularly deadly. Today the main dependable strategy utilized as a part of request to recognize sleep disorder is the polysomnography (PSG) which uses various recording such as electroencephalogram (EEG), electro-oculogram (EOG), electrocardiogram (ECG), body positions etc. [15]. One of the problems that hinder its use is the lack of sleep facilities, rather than the need for human experts, Therefore, it need an easier and cheaper solution. Henceforth, a strategy is proposed which forms brief span times of ECG signals.

Whatever is left of the paper is sorted out as takes after: Section II deals with materials and methods used in the paper and then section III deals with the conclusion of paper.

2. MATERIAL AND METHODS

2.1 Acquisition of data: In this paper, the experimental data of ECG signals related to sleep apnea and snoring is acquired from physionet.org. The beneath Fig.1 speaks to piece outline of the proposed framework.






Fig. 1. Block Diagram of Proposed System

2.2 Processing of ECG signals


In this study, a total of 145 patient’s ECG signals are taken from physio net database [16], out of which 105 patients belong to sleep apnea and 40 patients to snoring. The following Fig.2 represents ECG signals corresponding to normal, sleep apnea and snoring patients. To find out the pauses while breathing in sleep apnea and the occurrences of harsh/unpleasant sound caused by the vibration of soft palate in snoring, Tunable-Q factor wavelet transform (TQWT) [17] has been applied which is an adaptable and completely discrete wavelet transform. It is a valuable tool to analyse an oscillatory signal which accomplishes adaptability by changing its input parameters Q-factor Q that controls the quantity of oscillations of the wavelet. Now, the first step is to find out the TQWT parameters Q, R and J [18] which are obtained by verifying Parseval’s theorem. This theorem states that the energy of the wavelet coefficients must be equal to the energy of the original signal [18] as described in following Table 1. From table, it has been plainly delineated that for R value 5, it follows the Parseval’s theorem. Similarly values of Q and J are generated, i.e. 2 and 13 respectively.



(a)



(b)



(c)

Fig.2 ECG Signals related to normal (a), sleep apnea (b) and snoring patients (c)

The second step is based on the comparison of sub-bands of the apnea and normal patients as shown in the following Fig. 3. From this figure, it has been clearly depicted that (a) part has less amount of pauses at 0 values than (b) in which a patient is suffering from sleep apnea.

TABLE I
Comparison of Wavelet Coefficient Energy and Signal Energy

R Value	Wavelet Energy	Signal Energy
2	NaN	NaN
3	NaN	NaN
4	NaN	NaN
5	5.906875	5.906875
6	11.813750	5.906875
7	17.720625	5.906875
8	23.627500	5.906875
9	29.534375	5.906875
10	35.441250	5.906875
11	41.348125	5.906875
12	47.255000	5.906875

Likely, the snore values are computed in case of snoring patients as shown in the below Fig. 3. The histograms of the samples show variations in dispersion and steepness among normal, sleep apnea and snoring.




Fig. 3 Histograms of sub-bands (a,b) 4,8 respectively for normal patients, (c,d) 4 and 8 respectively for sleep apnea and (e,f) 4,nd 8 respectively for snoring patients

3. RESULTS AND DISCUSSION

In this paper an effort is made for the detection of sleep apnea and snoring by applying TQWT. For the observation purpose, ECG signals are collected from Physio net database. From older days to till now, PSG procedure is used to detect the sleep disorders but it is not possible to go through from the PSG procedure due to its high cost factor and need of more human specialists and unique labs. Moreover, it is also a time consuming and cumbersome process. Subsequently, there is a need of a more comfortable and less expensive technique to detect such types of disorders. Hence, ECG signals are used to detect the sleep apnea and snoring. The accuracy achieved in this method is 85% and can be improved with the addition of symptomatic guidelines.

REFERENCES

- [1] Sleep Disorder Overview, “www.neurologychannel.com”, Accessed: November 27, 2019.
- [2] G.Guimarae et al., “A method for automated temporal knowledge acquisition applied to sleep-related breathing disorders,” *artificial Intelligence in Medicine*, vol. 23, 2001, pp. 211-237.
- [3] D Liu, Z Pang, SR Lloyd, "A Neural Network Method for Detection of Obstructive Sleep Apnea and Narcolepsy Based on Pupil Size and EEG", *IEEE Transactions on Neural Networks*, vol. 19, Issue 2, 2008, pp. 308-318.
- [4] T. Lajnef, S. Chaibi, P. Ruby, P.E. Aguera, J.B. Eichenlaub, M. Samet, A. Kachouri, and K. Jerbi, “Learning machines and sleeping brains: Automatic sleep stage classification using decision-tree multi class support vector machines,” *journal of Neuroscience Methods*, 2015.
- [5] C.Oreilly and T. Nielsen, “Assessing eeg sleep spindle propagation. Part 1: Theory and proposed methodology,” *journal of Neuroscience Methods*, vol. 221, pp. 202-214, 2014.
- [6] V. Bajaj and R. B. Pachori, “Automatic classification of sleep stages based on the time-frequency image of eeg signals,” *Computer Methods and Programs in Biomedicine*, vol. 112, no. 3, pp. 320-328, 2013.
- [7] L. Fraiwan, K. Lweesy, N. Khasawneh, H. Wenz, and H. Dickhaus, “ Automated sleep stage identification system based on time-frequency analysis of a single eeg channel and random forest classifier,” *Computer Methods and Programs in Biomedicine*, 2011.
- [8] A Azarbarzin and Z Moussavi, “Snoring sounds variability as a signature of obstructive sleep apnea”, *Medical Engineering and Physics*, vol. 35 pp. 479–85, 2012
- [9] G Schlotthauer, L.E.D. Persia and Dand Milone,”Screening of obstructive sleep apnea with empirical mode decomposition of pulse oximetry”, *Medical Engineering and Physics*, vol. 36, pp 1074–80, 2014.

- [10] A R Hassan and M A Haque , “Computer-aided obstructive sleep apnea screening from single-lead electrocardiogram using statistical and spectral features and bootstrap aggregating”, *Biocybernetics and Biomedical Engineering*, vol. 36, pp 256–66, 2016.
- [11] A R Hassan and M A Haque, “Computer-aided sleep apnea diagnosis from single-lead electrocardiogram using dual tree complex wavelet transform and spectral features”, *Electrical and Electronic Engineering (ICEEE) 2015 Int. Conf.*, pp 1–4, 2015.
- [12] L Chen, Zhang Xand and C Song,” An automatic screening approach for obstructive sleep apnea diagnosis based on single-lead electrocardiogram”, *Automation Science and Engineering, IEEE Transactions*, vol. 12, pp 106–15, 2015.
- [13] C Song, K Liu, X Zhang, L Chen and X Xian, “An obstructive sleep apnea detection approach using a discriminative hidden markov model from ecg signals” , *Biomedical Engineering, IEEE Transactions*, vol. 63, pp 1–4, 2015.
- [14] M Smoleń, K Czopek and P Augustyniak, “Sleep Evaluation Device for Home-Care. In: Piętka E., Kawa J. (eds) *Information Technologies in Biomedicine*”, *Advances in Intelligent and Soft Computing*, vol 69, 2010.
- [15] A. Hachem, M. Ayache, L. El Khansa and A. Jezzini, "ECG classification for Sleep Apnea detection", 3rd Middle East Conference on Biomedical Engineering (MECBME), Beirut, pp.38-41,2016.
- [16] ECG database, “<https>” Accessed on October 6, 2011.
- [17] Ahnaf Rashik Hassan, “Computer-aided obstructive sleep apnea detection using normal inverse Gaussian parameters and adaptive boosting”, In *Biomedical Signal Processing and Control*, vol.29, pp.22, 2016.
- [18] Ivan Selesnick, “TQWT Toolbox Guide”, October 16, 2014.